Nano Energy

Volume 70, April 2020, 104420

Leonard Siebert a Niklas Wolff b Nicolai Ababii c Maik-lvo Terasa a Oleg Lupana c Alexander Vahl d Viola Duppel e Haoyi Qiu a Maik Tienken a Mattia Mirabelli a Victor Sontea c Franz Faupel d Lorenz Kienle b Rainer Adelung a

Chair for Functional Nanomaterials, Institute for Materials Science, Kiel University, Kaiserstr. 2, D-24143, Kiel, Germany

а

b

С

d

Chair for Synthesis and Real Structure, Institute for Materials Science, Kiel University, Kaiserstr. 2, D-24143, Kiel, Germany

Department of Microelectronics and Biomedical Engineering, Center for Nanotechnology and Nanosensors, Technical University of Moldova, 168 Stefan cel Mare Av., MD-2004, Chisinau, Republic of Moldova

Chair for Multicomponent Materials, Institute for Materials Science, Kiel University, Kaiserstr. 2, D-24143, Kiel, Germany

Nanochemistry, Max Planck Institute for Solid State Research, Heisenbergstr. 1, D-70569, Stuttgart, Germany

Facile fabrication of semiconducting oxide nanostructures by direct ink writing of readily available metal microparticles and their application as low power acetone gas sensors

DOI:10.1016/j.nanoen.2019.104420

Abstract

In this work, a facile two-step fabrication and characterization of printed acetone sensors based on mixed semiconducting metal oxides is introduced. The devices are fabricated by Direct Ink Writing metal microparticle (MP) stripes of commercially available pure iron and copper particles onto the surface of a glass substrate, forming a bridging multi-phase semiconducting oxide net by subsequent thermal annealing. The open, highly porous bridging structures consist of heterojunctions which are interconnected via non-planar CuO/Cu₂O/Cu nanowires and Fe₂O₃/Fe nanospikes. Morphological, vibrational, chemical and structural studies were performed to investigate the contact-forming Fe₂O₃-CuO nanostructures on the surface of the MPs. The power consumption and the gas sensing properties showed selectivity to acetone vapor at an operating temperature of around 300 °C with a high gas response of about 50% and the lowest operating power of around 0.26 µW to a concentration of 100 ppm of acetone vapor. The combination of the possibility of acetone vapor detection, the controllable size and geometry and their low power make these printed structures important candidates for next developments of accessible detection devices, as well as acetone vapor monitoring (even below 1 ppm). The printing of MPs in general paves the way for a new generation of printed different devices, even in "home-made" conditions, for a manifold of applications tailored by the composition and geometry of the printed MP stripes, enabled through the simplicity and versatility of the fabrication method.